



International Organization for Migration (IOM)

The UN Migration Agency

Reference No: AMM-2021-008

BID BULLETIN NO. 008-01

Supply and delivery of Digital X-Ray Machines (fixed/stationary) for Ministry of health centers

28-July-2021

The Bids Evaluation and Awards Committee of IOM Amman hereby issues Bid Bulletin No. 008-01 to announce new deadline for submissions: **15-August-2021, 14:30**, and to answer the following requests for clarifications:

- 1. It is mentioned that a floor mounted tube stand is required, and automatic collimator is required, but with the floor type configuration the automatic functions are not available because they need ceiling mount configuration. So please do you want floor mount type without automatic functions, or you want ceiling mount type with or without automatic collimator and functions.**

It is mentioned that a simple x-ray table with wheels is required, so please clarify if you want a full bucky table unit with grid and 2nd flat panel detector (in this case it does not come with wheels) or want only a simple stretcher on wheels and without any bucky device, grid and without 2nd flat panel detector.

IOM's official answer: we need a ceiling mount type with automatic collimator and functions, and simple stretcher on wheels, according to our room and work.

- 2. It is mentioned that "Pixel size/pitch, lower than 60%, required especially if the machine is single phase", please clarify what is the required in this specification (the pixel size or the DQE of the flat panel detector).**

IOM's official answer: the pixel size (100-200 microns) and the DQE up to 70% @0lp.

- 3. It is mentioned that "Number of detector: double/single", this also depends on if you want the full table bucky unit to be offered or not. So please clarify if we can offer the table bucky unit with 2nd flat panel detector as an optional item.**

IOM's official answer: the table with wheels does not need a second detector, but it can be included as optional item

- 4. It is mentioned that “Is the x-ray laser printer, if needed?”, please clarify if we can offer the digital film printer as an optional item.**

IOM’s official answer: we need an X-ray laser printer because the use of CD’s in digital film printer is not preferred.

- 5. We are assuming that the request is for one (1) Digital X-ray Machine?**

IOM’s official answer: The number of required X-ray machines is 12 (twelve).

- 6. Is this tender is open to international companies?**

IOM’s official answer: Yes, it is open to the International Companies. Please kindly check section 2 of ITB, eligible Bidders, for more details.

- 7. Please advise the authority that can issue the below requested certificate “Certification that Non-performance of contract did not occur within the last 3 years prior to application for evaluation based on all information on fully settled disputes or litigation”.**

IOM’s official answer: This is a confirmation on completed projects in the last 3 years, with a list giving full information (value of contract, contact information of referee...etc.)

- 8. Shall the price include sales tax, or it should be without sales tax.**

IOM’s official answer: Prices should be tax exempted, and as stated in clause 12.b, the price of the goods shall be quoted in DAP INCOTERMS 2010.

- 9. In case we can offer optional items, do we have to include the optional items prices in the required Bid Security or only the main system configuration must be taken into consideration.**

IOM’S official answer: Bid security is based on the value of items that make up an integral part of the x-ray, items that are part of the optional configuration should be excluded from the total value/bid security

- 10. Do we need to submit a price for annual service contract for the offered systems to be considered after the required two years warranty or it is not needed?**

IOM’S official answer: Not needed.

- 11. One of the required documents, to be submitted with the original bid, is the “Bid documents checklist (Form 7)” and this form is not included with the Bid Forms downloaded from the site (Section V), only five forms are available there, so please send us Form 7 and Form 6 or any other required forms and not included in Section V, if available.**

IOM’S official answer: Forms 6 (payment terms) and 7 (Bid documentation checklist) are now available on <https://jordan.iom.int/digital-x-ray-machine-fixedstationary-moh-centers>.

- 12. The required delivery time for the goods is 60 days approximately and this period is very short and difficult to be achieved due to the current worldwide situation of shipment and with this period**

companies will consider air freight instead of normal sea freight which will result in a significant price difference for the total quantity (12 sets).

IOM'S official answer: 60 is preferred is preferred by IOM, please indicate realistic time frame and suitable delivery modes for your company.

- 13. Kindly advise the installation sites of all 12 units (Names and contact details of each healthcare center), as our service team needs to visit the locations to check on the site readiness before quoting.**

IOM'S official answer: Room dimensions and photos are available at the following link:
https://iomint-my.sharepoint.com/:f/g/personal/galbader_iom_int/Elfp5id-QCVMxV04OmrNvMBUhTDryK38NZ44PhvKV_bhQ