



International Organization for Migration (IOM)

The UN Migration Agency

Reference No: AMM-2021-008

BID BULLETIN NO. 008-02

Supply and delivery of Digital X-Ray Machines (fixed/stationary) for Ministry of health centers

8-August-2021

The Bids Evaluation and Awards Committee of IOM Amman hereby issues Bid Bulletin No. 008-01 to answer the following requests for clarifications:

1. May we place our Bid in Euros, if different please advise.

IOM's official answer: As stated in clause 13.1 the prices in bid shall be quoted in national currency of Jordan (Jordanian Dinars).

2. Are the below points required for international companies?

Certificate of Registration from government agency in Jordan.

Confirmations of Tax Identification Number and certificate of taxpayer, issued by respective government agencies in Jordan; INT

IOM's official answer: The company who is not registered in Jordan shall provide package of the documents which proves duly registration/license issued by the respective government authority including proven tax registration/absence of the tax debts in the country where the company is registered and operating.

3. Can the following specifications be changed?

	Tender specification / Condition	Proposed changes	IOM's official answer
The collimator	Automatic/semiautomatic collimator with light source for centering and timer.	Automatic/semiautomatic/ Manual collimator with light source for centering and timer.	Manual can be added but priority will be given to Automatic/ semiautomatic options

Chest Bucky	Strong Bucky stand, well fixed on the floor o strong and automatic cassette adjustments and grid	Please clarify.	The Bucky is handled many times per day and the stand need to be physically strong and durable
The Digital X-ray Panel Detector	Pixel number and pixel size, higher pixel number, more than 2800x3000), and Pixel size/pitch, lower than 60%, required especially if the machine is single phase	Pixel number and pixel size, higher pixel number, more than 2,156 x 2,653, and Pixel size/pitch, lower than or equal 70%, required especially if the machine is single phase	This is important parameter of a detector, the lowest should be 2500X3000 The >60% mentioned for the panel detector is for DQE (Detective Quantum Efficiency). The higher is the better.
The Digital X-ray Panel Detector	Time to be ready for next image taking, shorter the better, less 10 secs	Time to be ready for next image taking, shorter the better, less or equal 11 secs	This is ok, not much difference. Preference might be given to the shorter period
The x-ray Operator's Computer Hardware and Software	Operator's (radiographer's) Screen/Monitor size minimum 21", RAM (8GB or more), processor (latest, i5 or above) and storage (1TB) or higher,	Operator's (radiographer's) Screen/Monitor size minimum 19", RAM (8GB or more), processor (latest, i5 or above) and storage (480 GB) or higher,	19" can be the minimum acceptable, but 21" is preferred Minimum storage can be 500GB, but preferably higher. Preference might be given to the initially indicated specs

IOM's official answer: Please note the specifications issued by IOM are approved by the Global Telerad team, and any changes made need to be approved by them. Please submit an offer that best matches these specifications, any equipment with alternative specifications could be included as an alternative option.